

## Linearity Test Report for Gage ID: Demo\_01

Gage ID	Demo_01	Part No. / Name	-	Coverage	99.73% (6 * Sigma)	Study Date	19/02/2016
Gage Type	Micrometer	Characteristic	-	Historical Pro. Std. Dev. (1 Sigma)	-	Tolerance	-
Resolution	0.001	Confidence	95%	Study By	-	-	-

Level = 5      Reading = 12

## Linearity Data

Level	1	2	3	4	5					
Reference Value	2.0	Bias	6.0	Bias	10.0					
1	2.7	0.7	5.1	1.1	5.8	-0.2	7.6	-0.4	9.1	-0.9
2	2.5	0.5	3.9	-0.1	5.7	-0.3	7.7	-0.3	9.3	-0.7
3	2.4	0.4	4.2	0.2	5.9	-0.1	7.8	-0.2	9.5	-0.5
4	2.5	0.5	5	1	5.9	-0.1	7.7	-0.3	9.3	-0.7
5	2.7	0.7	3.8	-0.2	6	0	7.8	-0.2	9.4	-0.6
6	2.3	0.3	3.9	-0.1	6.1	0.1	7.8	-0.2	9.5	-0.5
7	2.5	0.5	3.9	-0.1	6	0	7.8	-0.2	9.5	-0.5
8	2.5	0.5	3.9	-0.1	6.1	0.1	7.7	-0.3	9.5	-0.5
9	2.4	0.4	3.9	-0.1	6.4	0.4	7.8	-0.2	9.6	-0.4
10	2.4	0.4	4.	0	6.3	0.3	7.5	-0.5	9.2	-0.8
11	2.6	0.6	4.1	0.1	6	0	7.6	-0.4	9.3	-0.7
12	2.4	0.4	3.8	-0.2	6.1	0.1	7.7	-0.3	9.4	-0.6
Average	2.491667	0.491667	4.125000	0.125000	6.025000	0.025000	7.708333	-0.291667	9.383333	-0.616667

## Result

Statistic	Value
Regression Line: $y = ax + b$	
Slope (a)	-0.131667
Intercept (b)	0.736667
Repeatability Std. Dev. (s)	0.239540
Goodness of Fit (R-Sq)	71.43%
%EV/PV (Based on Process Variation)	NA
%EV/TV (Based on Tolerance)	NA
Hypothesis Test	
Significant t value	2.001717
$ ta $	12.042559
$ tb $	10.157519
Bias=0 line lies within the confidence bounds from 5.075978 & 6.068709	
Linearity Not Acceptable	

## Graphical Analysis

